

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/19/2024

Continuing Calib Time: 11:28 Initial Calibration Time(s): 16:06 02:09

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.43	4.43	4.36	4.50	0.00
Aroclor-1016-2	(2)	4.45	4.45	4.38	4.52	0.00
Aroclor-1016-3	(3)	4.50	4.50	4.43	4.57	0.00
Aroclor-1016-4	(4)	4.59	4.59	4.52	4.66	0.00
Aroclor-1016-5	(5)	4.88	4.88	4.81	4.95	0.00
Aroclor-1260-1	(1)	5.98	5.98	5.91	6.05	0.00
Aroclor-1260-2	(2)	6.24	6.24	6.17	6.31	0.00
Aroclor-1260-3	(3)	6.59	6.59	6.52	6.66	0.00
Aroclor-1260-4	(4)	6.81	6.81	6.74	6.88	0.01
Aroclor-1260-5	(5)	7.12	7.12	7.05	7.19	0.00
TCX		3.34	3.34	3.29	3.39	0.00
DCB		8.47	8.48	8.38	8.58	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/19/2024

Continuing Calib Time: 11:28 Initial Calibration Time(s): 16:06 02:09

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.73	3.73	3.66	3.80	0.00
Aroclor-1016-2	(2)	3.74	3.74	3.67	3.81	0.00
Aroclor-1016-3	(3)	3.90	3.90	3.83	3.97	0.00
Aroclor-1016-4	(4)	3.95	3.95	3.88	4.02	0.00
Aroclor-1016-5	(5)	4.15	4.15	4.08	4.22	0.00
Aroclor-1260-1	(1)	5.13	5.13	5.06	5.20	0.00
Aroclor-1260-2	(2)	5.32	5.32	5.25	5.39	0.00
Aroclor-1260-3	(3)	5.46	5.46	5.39	5.53	0.00
Aroclor-1260-4	(4)	5.92	5.92	5.85	5.99	0.00
Aroclor-1260-5	(5)	6.17	6.17	6.10	6.24	0.00
TCX		2.74	2.74	2.69	2.79	0.00
DCB		7.48	7.48	7.38	7.58	0.00