

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/19/2024

Continuing Calib Time: 13:37 Initial Calibration Time(s): 16:06 02:09

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.43	4.43	4.36	4.50	0.00
Aroclor-1242-2	(2)	4.45	4.45	4.38	4.52	0.00
Aroclor-1242-3	(3)	4.50	4.50	4.43	4.57	0.00
Aroclor-1242-4	(4)	4.60	4.60	4.53	4.67	0.01
Aroclor-1242-5	(5)	5.31	5.31	5.24	5.38	0.01
TCX		3.34	3.34	3.29	3.39	0.00
DCB		8.48	8.47	8.37	8.57	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/19/2024

Continuing Calib Time: 13:37 Initial Calibration Time(s): 16:06 02:09

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	3.73	3.73	3.66	3.80	0.00
Aroclor-1242-2	(2)	3.74	3.74	3.67	3.81	0.00
Aroclor-1242-3	(3)	3.90	3.90	3.83	3.97	0.00
Aroclor-1242-4	(4)	3.99	3.99	3.92	4.06	0.00
Aroclor-1242-5	(5)	4.48	4.48	4.41	4.55	0.00
TCX		2.74	2.74	2.69	2.79	0.00
DCB		7.48	7.48	7.38	7.58	0.00