

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/19/2024

Continuing Calib Time: 14:09 Initial Calibration Time(s): 16:06 02:09

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.25	5.24	5.17	5.31	0.00
Aroclor-1254-2	(2)	5.46	5.46	5.39	5.53	0.00
Aroclor-1254-3	(3)	5.81	5.81	5.74	5.88	0.00
Aroclor-1254-4	(4)	6.10	6.09	6.02	6.16	-0.01
Aroclor-1254-5	(5)	6.51	6.51	6.44	6.58	0.00
TCX		3.34	3.33	3.28	3.38	-0.01
DCB		8.47	8.47	8.37	8.57	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/19/2024

Continuing Calib Time: 14:09 Initial Calibration Time(s): 16:06 02:09

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	4.48	4.48	4.41	4.55	0.00
Aroclor-1254-2	(2)	4.63	4.63	4.56	4.70	0.00
Aroclor-1254-3	(3)	5.01	5.01	4.94	5.08	0.00
Aroclor-1254-4	(4)	5.23	5.23	5.16	5.30	0.00
Aroclor-1254-5	(5)	5.63	5.63	5.56	5.70	0.00
TCX		2.74	2.74	2.69	2.79	0.00
DCB		7.48	7.48	7.38	7.58	0.00