

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/19/2024

Continuing Calib Time: 21:32 Initial Calibration Time(s): 16:06 02:09

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------|-----|------------|-----------|-----------|------|------------|
|                |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1 | (1) | 5.24       | 5.24      | 5.17      | 5.31 | 0.00       |
| Aroclor-1254-2 | (2) | 5.46       | 5.46      | 5.39      | 5.53 | 0.00       |
| Aroclor-1254-3 | (3) | 5.81       | 5.81      | 5.74      | 5.88 | 0.00       |
| Aroclor-1254-4 | (4) | 6.09       | 6.09      | 6.02      | 6.16 | 0.00       |
| Aroclor-1254-5 | (5) | 6.51       | 6.51      | 6.44      | 6.58 | 0.00       |
| TCX            |     | 3.33       | 3.33      | 3.28      | 3.38 | 0.00       |
| DCB            |     | 8.47       | 8.47      | 8.37      | 8.57 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/19/2024

Continuing Calib Time: 21:32 Initial Calibration Time(s): 16:06 02:09

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND       |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------|-----|------------|-----------|-----------|------|------------|
|                |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1 | (1) | 4.48       | 4.48      | 4.41      | 4.55 | 0.00       |
| Aroclor-1254-2 | (2) | 4.63       | 4.63      | 4.56      | 4.70 | 0.00       |
| Aroclor-1254-3 | (3) | 5.00       | 5.01      | 4.94      | 5.08 | 0.01       |
| Aroclor-1254-4 | (4) | 5.23       | 5.23      | 5.16      | 5.30 | 0.00       |
| Aroclor-1254-5 | (5) | 5.63       | 5.63      | 5.56      | 5.70 | 0.00       |
| TCX            |     | 2.74       | 2.74      | 2.69      | 2.79 | 0.00       |
| DCB            |     | 7.48       | 7.48      | 7.38      | 7.58 | 0.00       |