

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG No.: P4886
Instrument ID: BNA_P Calibration Date(s): 11/08/2024 11/08/2024
Calibration Time(s): 18:53 22:58

LAB FILE ID:		RRFAL1 = BP022940.D			RRFAL2 = BP022941.D		RRFAL3 = BP022942.D	
		RRFAL4 = BP022943.D			RRFAL5 = BP022944.D		RRFAL6 = BP022945.D	
COMPOUND	RRFAL1	RRFAL2	RRFAL3	RRFAL4	RRFAL5	RRFAL6	RRF	% RSD
1,4-Dioxane	0.498	0.499	0.485	0.464	0.457		0.481	4.0
Benzaldehyde		1.019	1.045	0.911	0.710	0.569	0.851	24.2
Hexachloroethane	0.515	0.551	0.543	0.552	0.557		0.544	3.1
Nitrobenzene	0.370	0.382	0.418	0.416	0.412		0.400	5.5
Isophorone	0.643	0.679	0.754	0.757	0.739		0.714	7.1
2-Nitrophenol	0.155	0.166	0.183	0.184	0.188		0.175	8.1
2,4-Dimethylphenol	0.359	0.355	0.405	0.384	0.388		0.378	5.6
Bis(2-Chloroethoxy)methane	0.351	0.383	0.412	0.416	0.403		0.393	6.8
2,4-Dichlorophenol	0.311	0.314	0.366	0.369	0.361		0.344	8.4
Naphthalene	0.926	0.948	1.027	0.989	0.968		0.972	4.0
4-Chloroaniline		0.370	0.402	0.412	0.417	0.408	0.402	4.7
Hexachlorobutadiene	0.325	0.304	0.317	0.309	0.303		0.312	3.0
Phenol		1.401	1.516	1.585	1.631	1.588	1.544	5.8
Caprolactam		0.087	0.105	0.106	0.107	0.107	0.102	8.6
4-Chloro-3-methylphenol	0.324	0.348	0.396	0.394	0.390		0.370	8.8
2-Methylnaphthalene	0.696	0.709	0.764	0.753	0.747		0.734	4.0
Hexachlorocyclopentadiene		0.254	0.324	0.370	0.392	0.433	0.355	19.3
2,4,6-Trichlorophenol	0.364	0.383	0.422	0.441	0.446		0.411	8.9
2,4,5-Trichlorophenol	0.372	0.410	0.468	0.477	0.489		0.443	11.3
1,1-Biphenyl	1.181	1.267	1.337	1.329	1.294		1.282	4.9
2-Chloronaphthalene	0.959	1.041	1.090	1.075	1.072		1.048	5.0
2-Nitroaniline	0.233	0.273	0.309	0.327	0.329		0.294	14.0
Bis(2-Chloroethyl)ether		1.099	1.155	1.198	1.207	1.143	1.160	3.8
Dimethylphthalate	1.293	1.440	1.479	1.475	1.421		1.422	5.4
2,6-Dinitrotoluene	0.233	0.259	0.284	0.303	0.304		0.277	10.9
Acenaphthylene	1.332	1.492	1.633	1.575	1.570		1.520	7.7
3-Nitroaniline		0.220	0.253	0.250	0.238	0.231	0.238	5.7
Acenaphthene	0.998	1.095	1.147	1.150	1.129		1.104	5.7
2,4-Dinitrophenol		0.140	0.171	0.207	0.227	0.237	0.196	20.5
4-Nitrophenol		0.172	0.259	0.278	0.289	0.271	0.253	18.6
Dibenzofuran	1.590	1.696	1.793	1.750	1.699		1.705	4.5
2,4-Dinitrotoluene	0.366	0.424	0.470	0.472	0.471		0.441	10.5
Diethylphthalate	1.288	1.401	1.450	1.473	1.418		1.406	5.1
2-Chlorophenol	1.013	1.107	1.120	1.171	1.210		1.124	6.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG No.: P4886
Instrument ID: BNA_P Calibration Date(s): 11/08/2024 11/08/2024
Calibration Time(s): 18:53 22:58

LAB FILE ID:		RRFAL1 = BP022940.D			RRFAL2 = BP022941.D		RRFAL3 = BP022942.D	
		RRFAL4 = BP022943.D			RRFAL5 = BP022944.D		RRFAL6 = BP022945.D	
COMPOUND	RRFAL1	RRFAL2	RRFAL3	RRFAL4	RRFAL5	RRFAL6	RRF	% RSD
Fluorene	1.277	1.408	1.445	1.420	1.430		1.396	4.8
4-Chlorophenyl-phenylether	0.774	0.833	0.831	0.838	0.832		0.822	3.3
4-Nitroaniline		0.221	0.273	0.218	0.212	0.184	0.222	14.6
4,6-Dinitro-2-methylphenol		0.107	0.127	0.133	0.135	0.138	0.128	9.6
N-Nitrosodiphenylamine	0.418	0.447	0.492	0.489	0.464		0.462	6.6
1,2,4,5-Tetrachlorobenzene	0.636	0.660	0.703	0.695	0.682		0.675	4.1
4-Bromophenyl-phenylether	0.216	0.216	0.233	0.241	0.226		0.226	4.7
Hexachlorobenzene	0.279	0.280	0.297	0.288	0.282		0.285	2.7
Atrazine		0.211	0.227	0.232	0.224	0.223	0.223	3.5
Pentachlorophenol		0.152	0.172	0.181	0.183	0.197	0.177	9.4
2-Methylphenol		1.056	1.121	1.186	1.212	1.189	1.153	5.6
Phenanthrene	0.887	0.934	0.994	0.969	0.936		0.944	4.3
Anthracene	0.874	0.953	1.006	0.969	0.945		0.949	5.1
Carbazole		0.776	0.883	0.845	0.821	0.817	0.828	4.7
Di-n-butylphthalate	0.831	0.873	0.972	0.972	0.949		0.919	7.0
Fluoranthene	0.950	0.997	1.082	1.146	1.131		1.061	8.0
Pyrene	1.040	1.087	1.167	1.225	1.180		1.140	6.6
Butylbenzylphthalate	0.318	0.334	0.377	0.419	0.428		0.375	13.0
3,3-Dichlorobenzidine		0.422	0.482	0.456	0.450	0.435	0.449	5.1
2,2-oxybis(1-Chloropropane)		1.294	1.331	1.358	1.327	1.258	1.314	2.9
Benzo(a)anthracene	1.121	1.197	1.276	1.268	1.244		1.221	5.2
Chrysene	1.092	1.114	1.218	1.174	1.148		1.149	4.3
Bis(2-ethylhexyl)phthalate	0.429	0.496	0.545	0.589	0.608		0.533	13.6
Di-n-octyl phthalate		0.660	0.770	0.851	0.851	0.924	0.811	12.4
Benzo(b)fluoranthene	0.988	1.052	1.205	1.151	1.171		1.113	8.1
Benzo(k)fluoranthene	0.992	1.041	1.139	1.174	1.133		1.096	7.0
Benzo(a)pyrene	0.901	0.958	1.055	1.060	1.057		1.006	7.2
Indeno(1,2,3-cd)pyrene	1.372	1.346	1.464	1.417	1.411		1.402	3.2
Dibenzo(a,h)anthracene	1.116	1.102	1.183	1.141	1.150		1.138	2.7
Benzo(g,h,i)perylene	1.043	1.011	1.101	1.082	1.088		1.065	3.5
Acetophenone		1.921	2.046	2.134	2.098	2.025	2.045	4.0
2,3,4,6-Tetrachlorophenol	0.388	0.406	0.447	0.462	0.463		0.433	7.9
1,4-Dioxane-d8	0.387	0.411	0.440	0.419	0.423		0.416	4.7
Pyridine-d5		1.130	1.255	1.272	1.289	1.235	1.236	5.1
Phenol-d5		1.311	1.446	1.513	1.567	1.556	1.479	7.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG No.: P4886
Instrument ID: BNA_P Calibration Date(s): 11/08/2024 11/08/2024
Calibration Time(s): 18:53 22:58

LAB FILE ID:		RRFAL1 = BP022940.D			RRFAL2 = BP022941.D		RRFAL3 = BP022942.D	
		RRFAL4 = BP022943.D			RRFAL5 = BP022944.D		RRFAL6 = BP022945.D	
COMPOUND	RRFAL1	RRFAL2	RRFAL3	RRFAL4	RRFAL5	RRFAL6	RRF	% RSD
Bis-(2-Chloroethyl) ether-d8		0.842	0.876	0.892	0.889	0.846	0.869	2.7
2-Chlorophenol-d4	0.941	1.050	1.095	1.151	1.179		1.083	8.7
4-Methylphenol-d8		1.078	1.167	1.274	1.285	1.280	1.217	7.5
Nitrobenzene-d5	0.121	0.133	0.145	0.148	0.148		0.139	8.5
2-Nitrophenol-d4	0.142	0.153	0.171	0.181	0.182		0.166	10.5
2,4-Dichlorophenol-d3	0.302	0.331	0.360	0.373	0.372		0.348	8.9
4-Chloroaniline-d4		0.371	0.420	0.425	0.429	0.428	0.415	5.9
4-Methylphenol		1.166	1.252	1.329	1.330	1.289	1.273	5.4
Dimethylphthalate-d6	1.325	1.430	1.482	1.511	1.460		1.441	5.0
Acenaphthylene-d8	1.326	1.449	1.579	1.571	1.596		1.504	7.7
4-Nitrophenol-d4		0.166	0.214	0.229	0.241	0.233	0.217	13.9
Fluorene-d10	1.179	1.247	1.303	1.299	1.291		1.264	4.1
4,6-Dinitro-2-methylphenol		0.101	0.116	0.124	0.126	0.131	0.120	9.8
Anthracene-d10	0.773	0.805	0.917	0.882	0.857		0.847	6.8
Pyrene-d10	0.806	0.884	0.942	0.993	0.980		0.921	8.3
Benzo(a)pyrene-d12	0.839	0.881	1.016	1.012	1.016		0.953	9.0
N-Nitroso-di-n-propylamine	0.873	1.009	1.066	1.100	1.057		1.021	8.7

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1