

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886

EPA Sample No.: SSTD020730 Date Analyzed: 11/19/2024

Lab File ID: BP023219.D Time Analyzed: 15:58

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	71046	7.569	264596	10.32	208257	14.18
UPPER LIMIT	142092	8.069	529192	10.816	416514	14.681
LOWER LIMIT	35523	7.069	132298	9.816	104129	13.681
EPA SAMPLE NO.						
01 SVOC-GPC-BLANK	51880	7.58	176137	10.33	126737	14.19
02 SVOC-GPC2-BLANK	60018	7.57	202678	10.32	152370	14.18

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.