

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX043977.D
 Acq On : 25 Nov 2024 11:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 26 00:43:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	50.000	46.466	7.1	82	0.00
3 P	Chloromethane	50.000	46.775	6.5	86	0.00
4 C	Vinyl Chloride	50.000	46.231	7.5#	82	0.00
5 T	Bromomethane	50.000	51.566	-3.1	97	0.00
6 T	Chloroethane	50.000	48.469	3.1	87	0.00
7 T	Trichlorofluoromethane	50.000	44.839	10.3	80	0.00
8 T	Diethyl Ether	50.000	47.280	5.4	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.646	4.7	84	0.00
10 T	Methyl Iodide	50.000	47.795	4.4	85	0.00
11 T	Tert butyl alcohol	250.000	220.360	11.9	80	0.00
12 CM	1,1-Dichloroethene	50.000	46.080	7.8#	85	0.00
13 T	Acrolein	250.000	224.411	10.2	86	0.00
14 T	Allyl chloride	50.000	47.063	5.9	81	0.00
15 T	Acrylonitrile	250.000	239.244	4.3	83	0.00
16 T	Acetone	250.000	228.890	8.4	81	0.00
17 T	Carbon Disulfide	50.000	44.155	11.7	78	0.00
18 T	Methyl Acetate	50.000	46.531	6.9	83	0.00
19 T	Methyl tert-butyl Ether	50.000	48.373	3.3	86	0.00
20 T	Methylene Chloride	50.000	46.719	6.6	90	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.034	5.9	85	0.00
22 T	Diisopropyl ether	50.000	49.207	1.6	88	0.00
23 T	Vinyl Acetate	250.000	248.374	0.7	84	0.00
24 P	1,1-Dichloroethane	50.000	48.697	2.6	86	0.00
25 T	2-Butanone	250.000	236.327	5.5	81	0.00
26 T	2,2-Dichloropropane	50.000	47.384	5.2	83	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.444	3.1	86	0.00
28 T	Bromochloromethane	50.000	47.856	4.3	104	0.00
29 T	Tetrahydrofuran	250.000	232.975	6.8	82	0.00
30 C	Chloroform	50.000	49.337	1.3#	90	0.00
31 T	Cyclohexane	50.000	46.218	7.6	81	0.00
32 T	1,1,1-Trichloroethane	50.000	49.023	2.0	85	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.000	2.0	107	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	96	0.00
35 S	Dibromofluoromethane	50.000	50.333	-0.7	106	0.00
36 T	1,1-Dichloropropene	50.000	49.026	1.9	85	0.00
37 T	Ethyl Acetate	50.000	46.278	7.4	81	0.00
38 T	Carbon Tetrachloride	50.000	50.716	-1.4	88	0.00
39 T	Methylcyclohexane	50.000	47.843	4.3	81	0.00
40 TM	Benzene	50.000	48.936	2.1	87	0.00
41 T	Methacrylonitrile	50.000	48.098	3.8	80	0.00
42 TM	1,2-Dichloroethane	50.000	49.018	2.0	87	0.00
43 T	Isopropyl Acetate	50.000	48.245	3.5	84	0.00
44 TM	Trichloroethene	50.000	42.546	14.9	86	0.00
45 C	1,2-Dichloropropane	50.000	49.890	0.2#	88	0.00
46 T	Dibromomethane	50.000	50.387	-0.8	89	0.00
47 T	Bromodichloromethane	50.000	51.893	-3.8	89	0.00
48 T	Methyl methacrylate	50.000	48.244	3.5	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX112524\
 Data File : VX043977.D
 Acq On : 25 Nov 2024 11:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 26 00:43:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	869.321	13.1	77	0.00
50 S	Toluene-d8	50.000	49.805	0.4	107	0.00
51 T	4-Methyl-2-Pentanone	250.000	238.823	4.5	81	0.00
52 CM	Toluene	50.000	50.022	-0.0#	87	0.00
53 T	t-1,3-Dichloropropene	50.000	51.600	-3.2	86	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.373	-0.7	85	0.00
55 T	1,1,2-Trichloroethane	50.000	49.075	1.8	87	0.00
56 T	Ethyl methacrylate	50.000	49.531	0.9	83	0.00
57 T	1,3-Dichloropropane	50.000	49.872	0.3	89	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	226.771	9.3	83	0.00
59 T	2-Hexanone	250.000	241.619	3.4	80	0.00
60 T	Dibromochloromethane	50.000	51.814	-3.6	87	0.00
61 T	1,2-Dibromoethane	50.000	49.645	0.7	86	0.00
62 S	4-Bromofluorobenzene	50.000	50.783	-1.6	106	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	93	0.00
64 T	Tetrachloroethene	50.000	50.133	-0.3	88	0.00
65 PM	Chlorobenzene	50.000	49.197	1.6	87	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.189	-4.4	87	0.00
67 C	Ethyl Benzene	50.000	50.561	-1.1#	86	0.00
68 T	m/p-Xylenes	100.000	101.696	-1.7	86	0.00
69 T	o-Xylene	50.000	50.067	-0.1	83	0.00
70 T	Styrene	50.000	52.353	-4.7	86	0.00
71 P	Bromoform	50.000	46.066	7.9	81	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	48.163	3.7	85	0.00
74 T	N-amyl acetate	50.000	48.361	3.3	82	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.323	5.4	85	0.00
76 T	1,2,3-Trichloropropane	50.000	47.389	5.2	83	0.00
77 T	Bromobenzene	50.000	48.374	3.3	86	0.00
78 T	n-propylbenzene	50.000	49.892	0.2	86	0.00
79 T	2-Chlorotoluene	50.000	47.802	4.4	86	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.689	0.6	86	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.951	8.1	79	0.00
82 T	4-Chlorotoluene	50.000	48.585	2.8	85	0.00
83 T	tert-Butylbenzene	50.000	49.242	1.5	85	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.987	0.0	87	0.00
85 T	sec-Butylbenzene	50.000	49.775	0.5	85	0.00
86 T	p-Isopropyltoluene	50.000	51.764	-3.5	87	0.00
87 T	1,3-Dichlorobenzene	50.000	48.622	2.8	87	0.00
88 T	1,4-Dichlorobenzene	50.000	47.493	5.0	86	0.00
89 T	n-Butylbenzene	50.000	50.989	-2.0	83	0.00
90 T	Hexachloroethane	50.000	48.643	2.7	83	0.00
91 T	1,2-Dichlorobenzene	50.000	48.588	2.8	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.127	7.7	81	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.992	4.0	83	0.00
94 T	Hexachlorobutadiene	50.000	48.575	2.8	88	0.00
95 T	Naphthalene	50.000	49.325	1.3	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043977.D
Acq On : 25 Nov 2024 11:27
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 26 00:43:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.584	2.8	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6