



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043976.D
Acq On : 25 Nov 2024 10:59
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Title : SW846 8260
Last Update : Fri Nov 22 03:45:52 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	6761	PASS
75	95	30	60	58.2	17633	PASS
95	95	100	100	100.0	30288	PASS
96	95	5	9	7.5	2258	PASS
173	174	0.00	2	0.8	174	PASS
174	95	50	100	73.6	22298	PASS
175	174	5	9	7.5	1678	PASS
176	174	95	101	95.6	21315	PASS
177	176	5	9	6.9	1462	PASS