

Report of Analysis

Client:	Chemtech Consulting Group		Date Collected:	
Project:	Weekly Storage Blanks		Date Received:	
Client Sample ID:	VX1125WBS01		SDG No.:	P4886
Lab Sample ID:	VX1125WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC- Chemtech Full
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX043980.D	1		11/25/24 12:41	VX112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	17.0		0.21	1.00	ug/L
74-87-3	Chloromethane	18.3		0.35	1.00	ug/L
75-01-4	Vinyl Chloride	17.5		0.34	1.00	ug/L
141-78-6	Ethyl Acetate	17.0		0.38	1.00	ug/L
74-83-9	Bromomethane	20.6		1.40	5.00	ug/L
75-00-3	Chloroethane	19.4		0.56	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.0		0.34	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	17.5		0.25	1.00	ug/L
75-65-0	Tert butyl alcohol	85.3		5.60	25.0	ug/L
75-35-4	1,1-Dichloroethene	17.2		0.26	1.00	ug/L
107-02-8	Acrolein	79.5		6.70	25.0	ug/L
107-13-1	Acrylonitrile	90.5		0.90	5.00	ug/L
67-64-1	Acetone	87.8		1.40	5.00	ug/L
75-15-0	Carbon Disulfide	15.0		0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.0		0.16	1.00	ug/L
79-20-9	Methyl Acetate	18.3		0.60	1.00	ug/L
75-09-2	Methylene Chloride	17.8		0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	17.9		0.25	1.00	ug/L
108-05-4	Vinyl Acetate	92.9		0.71	5.00	ug/L
75-34-3	1,1-Dichloroethane	19.3		0.23	1.00	ug/L
110-82-7	Cyclohexane	17.9		1.60	5.00	ug/L
78-93-3	2-Butanone	91.3		1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	17.8		0.25	1.00	ug/L
594-20-7	2,2-Dichloropropane	18.5		0.31	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.7		0.25	1.00	ug/L
74-97-5	Bromochloromethane	21.3		0.18	1.00	ug/L
67-66-3	Chloroform	19.5		0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.3		0.19	1.00	ug/L
108-87-2	Methylcyclohexane	17.0		0.19	1.00	ug/L
563-58-6	1,1-Dichloropropene	18.8		0.18	1.00	ug/L

Report of Analysis

Client:	Chemtech Consulting Group		Date Collected:	
Project:	Weekly Storage Blanks		Date Received:	
Client Sample ID:	VX1125WBS01		SDG No.:	P4886
Lab Sample ID:	VX1125WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC- Chemtech Full
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX043980.D	1		11/25/24 12:41	VX112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
71-43-2	Benzene	18.8		0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.1		0.24	1.00	ug/L
79-01-6	Trichloroethene	16.1		0.32	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.8		0.19	1.00	ug/L
74-95-3	Dibromomethane	19.1		0.23	1.00	ug/L
75-27-4	Bromodichloromethane	18.8		0.24	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	92.4		0.75	5.00	ug/L
108-88-3	Toluene	19.1		0.18	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	18.2		0.21	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.6		0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	18.7		0.21	1.00	ug/L
142-28-9	1,3-Dichloropropane	19.5		0.17	1.00	ug/L
110-75-8	2-Chloroethyl Vinyl ether	76.4		1.80	5.00	ug/L
591-78-6	2-Hexanone	90.6		1.10	5.00	ug/L
124-48-1	Dibromochloromethane	17.9		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	18.5		0.16	1.00	ug/L
127-18-4	Tetrachloroethene	18.1		0.25	1.00	ug/L
108-90-7	Chlorobenzene	18.6		0.13	1.00	ug/L
630-20-6	1,1,1,2-Tetrachloroethane	19.1		0.21	1.00	ug/L
67-72-1	Hexachloroethane	16.7		0	0	ug/L
100-41-4	Ethyl Benzene	18.6		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	37.3		0.31	2.00	ug/L
95-47-6	o-Xylene	18.7		0.14	1.00	ug/L
100-42-5	Styrene	18.6		0.16	1.00	ug/L
75-25-2	Bromoform	16.0		0.21	1.00	ug/L
98-82-8	Isopropylbenzene	19.0		0.13	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.1		0.27	1.00	ug/L
96-18-4	1,2,3-Trichloropropane	19.0		0.39	1.00	ug/L
108-86-1	Bromobenzene	19.0		0.26	1.00	ug/L
103-65-1	n-propylbenzene	19.0		0.14	1.00	ug/L
95-49-8	2-Chlorotoluene	19.1		0.16	1.00	ug/L

Report of Analysis

Client:	Chemtech Consulting Group		Date Collected:	
Project:	Weekly Storage Blanks		Date Received:	
Client Sample ID:	VX1125WBS01		SDG No.:	P4886
Lab Sample ID:	VX1125WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC- Chemtech Full
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX043980.D	1		11/25/24 12:41	VX112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
108-67-8	1,3,5-Trimethylbenzene	19.8		0.18	1.00	ug/L
106-43-4	4-Chlorotoluene	19.1		0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	19.2		0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	19.8		0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	18.3		0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	19.5		0.15	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.6		0.24	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.5		0.27	1.00	ug/L
104-51-8	n-Butylbenzene	18.6		0.22	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.9		0.19	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.5		0.46	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.2		0.42	1.00	ug/L
87-68-3	Hexachlorobutadiene	18.8		0.31	1.00	ug/L
91-20-3	Naphthalene	17.7		0.59	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.9		0.51	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.6		74 - 125	101%	SPK: 50
1868-53-7	Dibromofluoromethane	49.4		75 - 124	99%	SPK: 50
2037-26-5	Toluene-d8	48.8		86 - 113	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.3		77 - 121	95%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	129000	5.544			
540-36-3	1,4-Difluorobenzene	233000	6.757			
3114-55-4	Chlorobenzene-d5	201000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	89400	12.024			



Client:	Chemtech Consulting Group			Date Collected:	
Project:	Weekly Storage Blanks			Date Received:	
Client Sample ID:	VX1125WBS01			SDG No.:	P4886
Lab Sample ID:	VX1125WBS01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOC- Chemtech Full
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX043980.D	1		11/25/24 12:41	VX112524

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

A = Aldol-Condensation Reaction Products