

Surrogate Summary

SDG No.: **P4886**

Client: **Chemtech Consulting Group**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4886-01	Storage-Blank-SOIL-REF-6	1,2-Dichloroethane-d4	50	51.4	103	74	125
		Dibromofluoromethane	50	47.2	94	75	124
		Toluene-d8	50	49.7	99	86	113
		4-Bromofluorobenzene	50	47.8	96	77	121
P4886-02	Storage-Blank-WATER-REF-5	1,2-Dichloroethane-d4	50	49.3	99	74	125
		Dibromofluoromethane	50	46.6	93	75	124
		Toluene-d8	50	49.8	100	86	113
		4-Bromofluorobenzene	50	47.6	95	77	121
P4886-03	Storage-Blank-WATER-REF-4	1,2-Dichloroethane-d4	50	50.2	100	74	125
		Dibromofluoromethane	50	46.5	93	75	124
		Toluene-d8	50	49.4	99	86	113
		4-Bromofluorobenzene	50	47.4	95	77	121
P4886-04	Storage-Blank-SAMPLE-MANAGEMENT	1,2-Dichloroethane-d4	50	50.2	100	74	125
		Dibromofluoromethane	50	47.5	95	75	124
		Toluene-d8	50	50.4	101	86	113
		4-Bromofluorobenzene	50	49.8	100	77	121
VX1125WBL01	VX1125WBL01	1,2-Dichloroethane-d4	50	52.0	104	74	125
		Dibromofluoromethane	50	47.2	94	75	124
		Toluene-d8	50	51.6	103	86	113
		4-Bromofluorobenzene	50	50.7	101	77	121
VX1125WBS01	VX1125WBS01	1,2-Dichloroethane-d4	50	50.6	101	74	125
		Dibromofluoromethane	50	49.4	99	75	124
		Toluene-d8	50	48.8	98	86	113
		4-Bromofluorobenzene	50	47.3	95	77	121
VX1125WBSD0	VX1125WBSD01	1,2-Dichloroethane-d4	50	49.6	99	74	125
		Dibromofluoromethane	50	48.7	97	75	124
		Toluene-d8	50	48.8	98	86	113
		4-Bromofluorobenzene	50	48.3	97	77	121