

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: P4886 SAS No.: P4886 SDG NO.: P4886
 Lab File ID: VX043977.D Date Analyzed: 11/25/2024
 Instrument ID: MSVOA_X Time Analyzed: 11:27
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	94886	12.018				
UPPER LIMIT	189772	12.518				
LOWER LIMIT	47443	11.518				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	78448	12.02				
Storage-Blank-WATER-REF-5	84391	12.02				
Storage-Blank-WATER-REF-4	75443	12.02				
Storage-Blank-SAMPLE-MANAGEMENT	82714	12.02				
VX1125WBL01	90683	12.02				
VX1125WBS01	89446	12.02				
VX1125WBSD01	89463	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.