

CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4888 SAS No.: P4888 SDG NO.: P4888

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 15:29 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.93	5.92	5.82	6.02	-0.01
Aroclor-1016-2	(2)	5.95	5.94	5.84	6.04	-0.01
Aroclor-1016-3	(3)	6.01	6.01	5.91	6.11	0.00
Aroclor-1016-4	(4)	6.11	6.11	6.01	6.21	0.00
Aroclor-1016-5	(5)	6.40	6.40	6.30	6.50	0.00
Aroclor-1260-1	(1)	7.53	7.52	7.42	7.62	-0.01
Aroclor-1260-2	(2)	7.78	7.78	7.68	7.88	0.00
Aroclor-1260-3	(3)	8.14	8.14	8.04	8.24	0.00
Aroclor-1260-4	(4)	8.38	8.38	8.28	8.48	0.00
Aroclor-1260-5	(5)	8.72	8.71	8.61	8.81	-0.01
Tetrachloro-m-xylene		4.76	4.76	4.66	4.86	0.00
Decachlorobiphenyl		10.68	10.67	10.57	10.77	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4888 SAS No.: P4888 SDG NO.: P4888

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 15:29 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.16	5.16	5.06	5.26	0.00
Aroclor-1016-2	(2)	5.18	5.18	5.08	5.28	0.00
Aroclor-1016-3	(3)	5.36	5.36	5.26	5.46	0.00
Aroclor-1016-4	(4)	5.40	5.40	5.30	5.50	0.00
Aroclor-1016-5	(5)	5.62	5.62	5.52	5.72	0.01
Aroclor-1260-1	(1)	6.66	6.66	6.56	6.76	0.00
Aroclor-1260-2	(2)	6.85	6.85	6.75	6.95	0.01
Aroclor-1260-3	(3)	7.00	7.01	6.91	7.11	0.01
Aroclor-1260-4	(4)	7.48	7.48	7.38	7.58	0.00
Aroclor-1260-5	(5)	7.72	7.72	7.62	7.82	0.00
Tetrachloro-m-xylene		4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl		9.21	9.21	9.11	9.31	0.00