

CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4888 SAS No.: P4888 SDG NO.: P4888

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/19/2024

Lab Sample No.: AR1660CCC500 Data File : PP068614.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.926	5.821	6.021	472.020	500.000	-5.6
Aroclor-1016-2	5.947	5.844	6.044	493.840	500.000	-1.2
Aroclor-1016-3	6.011	5.908	6.108	514.490	500.000	2.9
Aroclor-1016-4	6.109	6.005	6.205	518.260	500.000	3.7
Aroclor-1016-5	6.404	6.299	6.499	470.310	500.000	-5.9
Aroclor-1260-1	7.527	7.423	7.623	488.300	500.000	-2.3
Aroclor-1260-2	7.780	7.676	7.876	469.560	500.000	-6.1
Aroclor-1260-3	8.142	8.038	8.238	446.600	500.000	-10.7
Aroclor-1260-4	8.380	8.276	8.476	435.780	500.000	-12.8
Aroclor-1260-5	8.717	8.613	8.813	439.380	500.000	-12.1
Decachlorobiphenyl	10.675	10.572	10.772	45.710	50.000	-8.6
Tetrachloro-m-xylene	4.761	4.657	4.857	49.560	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4888 SAS No.: P4888 SDG NO.: P4888

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/19/2024

Lab Sample No.: AR1660CCC500 Data File : PP068614.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.156	5.058	5.258	501.490	500.000	0.3
Aroclor-1016-2	5.176	5.079	5.279	490.400	500.000	-1.9
Aroclor-1016-3	5.356	5.259	5.459	452.510	500.000	-9.5
Aroclor-1016-4	5.396	5.298	5.498	459.270	500.000	-8.1
Aroclor-1016-5	5.615	5.517	5.717	511.580	500.000	2.3
Aroclor-1260-1	6.658	6.560	6.760	492.620	500.000	-1.5
Aroclor-1260-2	6.845	6.747	6.947	463.600	500.000	-7.3
Aroclor-1260-3	7.002	6.905	7.105	472.300	500.000	-5.5
Aroclor-1260-4	7.477	7.378	7.578	478.300	500.000	-4.3
Aroclor-1260-5	7.715	7.617	7.817	470.590	500.000	-5.9
Decachlorobiphenyl	9.207	9.110	9.310	46.580	50.000	-6.8
Tetrachloro-m-xylene	4.051	3.951	4.151	50.940	50.000	1.9