

Surrogate Summary

SDG No.: **P4888**

Client: **BSI America Professional Services Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP068360.D	PIBLK-PP068360.D	Tetrachloro-m-xylene	1	20	21.0	105		60	140
		Decachlorobiphenyl	1	20	23.4	117		60	140
		Tetrachloro-m-xylene	2	20	21.6	108		60	140
		Decachlorobiphenyl	2	20	23.6	118		60	140
I.BLK-PP068606.D	PIBLK-PP068606.D	Tetrachloro-m-xylene	1	20	21.5	107		60	140
		Decachlorobiphenyl	1	20	22.1	111		60	140
		Tetrachloro-m-xylene	2	20	22.0	110		60	140
		Decachlorobiphenyl	2	20	22.2	111		60	140
PB165084BL	PB165084BL	Tetrachloro-m-xylene	1	20	20.5	102		32	144
		Decachlorobiphenyl	1	20	20.7	104		32	175
		Tetrachloro-m-xylene	2	20	20.2	101		32	144
		Decachlorobiphenyl	2	20	21.0	105		32	175
PB165084BS	PB165084BS	Tetrachloro-m-xylene	1	20	20.3	102		32	144
		Decachlorobiphenyl	1	20	20.5	102		32	175
		Tetrachloro-m-xylene	2	20	20.5	103		32	144
		Decachlorobiphenyl	2	20	20.9	104		32	175
P4888-01	01A-01B-01C	Tetrachloro-m-xylene	1	20	20.9	105		32	144
		Decachlorobiphenyl	1	20	22.2	111		32	175
		Tetrachloro-m-xylene	2	20	22.4	112		32	144
		Decachlorobiphenyl	2	20	22.0	110		32	175
P4888-02	02A-02B-02C	Tetrachloro-m-xylene	1	20	20.9	104		32	144
		Decachlorobiphenyl	1	20	21.6	108		32	175
		Tetrachloro-m-xylene	2	20	21.4	107		32	144
		Decachlorobiphenyl	2	20	21.6	108		32	175
I.BLK-PP068615.D	PIBLK-PP068615.D	Tetrachloro-m-xylene	1	20	20.9	105		60	140
		Decachlorobiphenyl	1	20	22.4	112		60	140
		Tetrachloro-m-xylene	2	20	21.9	110		60	140
		Decachlorobiphenyl	2	20	22.6	113		60	140