



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture
Project: Amtrak Sawtooth Bridges 2024

SDG No.: P4892
RunNo.: LB133491

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	11/18/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	11/18/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/18/2024

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: P4892

Project: Amtrak Sawtooth Bridges 2024

RunNo.: LB133497

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	11/18/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/18/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/18/2024
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	11/18/2024
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	11/18/2024

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: P4892

Project: Amtrak Sawtooth Bridges 2024

RunNo.: LB133500

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Corrosivity		pH	7.01	7	100	90-110	11/18/2024
Sample ID:	CCV1						
Corrosivity		pH	2.01	2.00	101	90-110	11/18/2024
Sample ID:	CCV2						
Corrosivity		pH	12.02	12.00	100	90-110	11/18/2024

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: P4892

Project: Amtrak Sawtooth Bridges 2024

RunNo.: LB133501

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.094	0.099	95	85-115	11/18/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	11/18/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	11/18/2024
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	11/18/2024