

rteres

Instrument :

FID_C

LabSampleId :

20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/25/2024

Supervised By :Ankita Jodhani 11/25/2024

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC11222
Data File : FC067858.D
Signal(s) : FID1A.ch
Acq On : 22 Nov 2024 16:57
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111424.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.287	3.259	3.339	BB	365200	3049441	86.30%	4.725%
2	4.349	4.315	4.412	BB	341350	3156009	89.31%	4.890%
3	5.926	5.894	6.027	BB	367311	3464972	98.06%	5.369%
4	6.362	6.327	6.427	BB	315158	3208299	90.79%	4.971%
5	6.980	6.939	7.092	BB	331615	3360369	95.10%	5.207%
6	8.154	8.095	8.224	BB	298037	3192198	90.34%	4.946%
7	9.752	9.709	9.830	BB	294539	3298674	93.35%	5.111%
8	11.191	11.094	11.274	BB	286640	3315554	93.83%	5.137%
9	11.502	11.452	11.567	BB	333375	3533591	100.00%	5.475%
10	12.498	12.459	12.555	BB	260863	3104156	87.85%	4.810%
11	12.940	12.880	13.002	BB	213763	2632448	74.50%	4.079%
12	13.109	13.069	13.177	BB	241250	3011752	85.23%	4.666%
13	13.694	13.662	13.762	VB	229389	2962357	83.83%	4.590%
14	14.795	14.727	14.870	BB	217012	2901804	82.12%	4.496%
15	15.814	15.754	15.879	BB	218859	2806749	79.43%	4.349%
16	16.763	16.620	16.839	BB	210338	2846652	80.56%	4.411%
17	17.649	17.579	17.707	BB	211588	3056177	86.49%	4.735%
18	18.480	18.409	18.552	BB	204299	3038765	86.00%	4.708%
19	19.262	19.204	19.329	BB	181039	2593871	73.41%	4.019%
20	20.000	19.945	20.035	BV	144964	2146608	60.75%	3.326%
21	20.724	20.660	20.795	BV	108307	1926790	54.53%	2.985%
22	21.612	21.529	21.739	BB	77348	1933778	54.73%	2.996%
Sum of corrected areas:							64541016	

Aliphatic EPH 111424.M Fri Nov 22 23:58:35 2024