

Continuing Calibration Report for SequenceID : FC112124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067839.D

Aliphatic C9-C12	9594932.000	60.000	3.186	6.463	159915.533	165791.359	3.544
Aliphatic C12-C16	6660895.000	40.000	6.464	9.851	166522.375	170384.422	2.267
Aliphatic C16-C21	9836167.000	60.000	9.852	13.208	163936.117	169621.553	3.352
Aliphatic C21-C28	12842540.000	80.000	13.209	16.863	160531.750	157280.060	-2.067
Aliphatic C28-C40	15281804.000	120.000	16.864	21.710	127348.367	121111.336	-5.150
Aliphatic EPH	54216338.000	360.000	3.186	21.710	150600.939	150155.324	-0.297