

Continuing Calibration Report for SequenceID : FC112124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067845.D

Aliphatic C9-C12	9334629.000	60.000	3.186	6.463	155577.150	165791.359	6.161
Aliphatic C12-C16	6473028.000	40.000	6.464	9.851	161825.700	170384.422	5.023
Aliphatic C16-C21	9664465.000	60.000	9.852	13.208	161074.417	169621.553	5.039
Aliphatic C21-C28	12821556.000	80.000	13.209	16.863	160269.450	157280.060	-1.901
Aliphatic C28-C40	14816190.000	120.000	16.864	21.710	123468.250	121111.336	-1.946
Aliphatic EPH	53109868.000	360.000	3.186	21.710	147527.411	150155.324	1.750