

Continuing Calibration Report for SequenceID : FC112224AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067848.D

Aliphatic C9-C12	8980746.000	60.000	3.186	6.463	149679.100	165791.359	9.718
Aliphatic C12-C16	6170368.000	40.000	6.464	9.851	154259.200	170384.422	9.464
Aliphatic C16-C21	9033880.000	60.000	9.852	13.208	150564.667	169621.553	11.235
Aliphatic C21-C28	11143774.000	80.000	13.209	16.863	139297.175	157280.060	11.434
Aliphatic C28-C40	14283251.000	120.000	16.864	21.709	119027.092	121111.336	1.721
Aliphatic EPH	49612019.000	360.000	3.186	21.709	137811.164	150155.324	8.221