

Continuing Calibration Report for SequenceID : FC112224AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067858.D

Aliphatic C9-C12	9413749.000	60.000	3.186	6.463	156895.817	165791.359	5.366
Aliphatic C12-C16	6490872.000	40.000	6.464	9.851	162271.800	170384.422	4.761
Aliphatic C16-C21	9422898.000	60.000	9.852	13.208	157048.300	169621.553	7.413
Aliphatic C21-C28	11509924.000	80.000	13.209	16.863	143874.050	157280.060	8.524
Aliphatic C28-C40	14695989.000	120.000	16.864	21.709	122466.575	121111.336	-1.119
Aliphatic EPH	51533432.000	360.000	3.186	21.709	143148.422	150155.324	4.666