

Continuing Calibration Report for SequenceID : FD112124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048790.D

Aromatic C10-C12	8107483.000	40.000	4.088	5.805	202687.075	192017.544	-5.557
Aromatic C12-C16	12127218.000	60.000	5.806	8.411	202120.300	192458.977	-5.020
Aromatic C16-C21	15266390.000	80.000	8.412	12.676	190829.875	185804.816	-2.704
Aromatic C21-C36	29247862.000	180.000	12.677	18.084	162488.122	157231.056	-3.344
Aromatic EPH	64748953.000	360.000	4.088	18.084	179858.203	173317.266	-3.774