

Continuing Calibration Report for SequenceID : FD112124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048796.D

Aromatic C10-C12	7963599.000	40.000	4.088	5.805	199089.975	192017.544	-3.683
Aromatic C12-C16	11946040.000	60.000	5.806	8.411	199100.667	192458.977	-3.451
Aromatic C16-C21	15031877.000	80.000	8.412	12.676	187898.463	185804.816	-1.127
Aromatic C21-C36	29031182.000	180.000	12.677	18.084	161284.344	157231.056	-2.578
Aromatic EPH	63972698.000	360.000	4.088	18.084	177701.939	173317.266	-2.530