

Continuing Calibration Report for SequenceID : FD112224AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048809.D

Aromatic C10-C12	7879801.000	40.000	4.088	5.804	196995.025	192017.544	-2.592
Aromatic C12-C16	11762602.000	60.000	5.805	8.412	196043.367	192458.977	-1.862
Aromatic C16-C21	14657115.000	80.000	8.413	12.677	183213.938	185804.816	1.394
Aromatic C21-C36	28600750.000	180.000	12.678	18.087	158893.056	157231.056	-1.057
Aromatic EPH	62900268.000	360.000	4.088	18.087	174722.967	173317.266	-0.811