

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4892 **SAS No :** P4892 **SDG No:** P4892
Sample No : P4892-02MS **Datafile:** FC067855.D
Client ID : WB-310-BOTMS

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	11.6	0	8.97	77		(40-140)
Aliphatic C12-C16	7.7	0	8.25	106		(40-140)
Aliphatic C16-C21	11.6	0	13.1	112		(40-140)
Aliphatic C21-C28	15.5	0	17.1	110		(40-140)
Aliphatic C28-C40	23.2	0	28.6	123		(40-140)