

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4892 **SAS No :** P4892 **SDG No:** P4892
Sample No : PB165151BS **Datafile:** FC067841.D
Client ID : PB165151BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	164	55		(40-140)
Aliphatic C12-C16	200	143	72		(40-140)
Aliphatic C16-C21	300	241	80		(40-140)
Aliphatic C21-C28	400	325	81		(40-140)
Aliphatic C28-C40	600	554	92		(40-140)