

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4892 **SAS No :** P4892 **SDG No:** P4892
Sample No : PB165188BS **Datafile:** FC067850.D
Client ID : PB165188BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10.0	5.81	58		(40-140)
Aliphatic C12-C16	6.7	5.08	76		(40-140)
Aliphatic C16-C21	10.0	8.29	83		(40-140)
Aliphatic C21-C28	13.3	11.0	82		(40-140)
Aliphatic C28-C40	20.0	18.4	92		(40-140)