

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4892 **SAS No :** P4892 **SDG No:** P4892
Sample No : PB165151BS **Datafile:** FD048792.D
Client ID : PB165151BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	179	90		(40-140)
Aromatic C12-C16	300	321	107		(40-140)
Aromatic C16-C21	500	593	118		(40-140)
Aromatic C21-C36	800	940	117		(40-140)