

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 20:36 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.53	5.53	5.43	5.63	0.00
Aroclor-1248-2 (2)	5.80	5.80	5.70	5.90	0.00
Aroclor-1248-3 (3)	6.01	6.00	5.90	6.10	-0.01
Aroclor-1248-4 (4)	6.41	6.41	6.31	6.51	0.00
Aroclor-1248-5 (5)	6.45	6.45	6.35	6.55	0.00
Tetrachloro-m-xylene	4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl	10.10	10.09	9.99	10.19	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 20:36 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1248-2 (2)	4.96	4.96	4.86	5.06	0.01
Aroclor-1248-3 (3)	5.00	5.00	4.90	5.10	0.00
Aroclor-1248-4 (4)	5.17	5.17	5.07	5.27	0.00
Aroclor-1248-5 (5)	5.56	5.56	5.46	5.66	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.63	8.64	8.54	8.74	0.01