

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 20:53 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.38	6.38	6.28	6.48	0.00
Aroclor-1254-2	(2)	6.60	6.60	6.50	6.70	0.00
Aroclor-1254-3	(3)	6.97	6.97	6.87	7.07	0.00
Aroclor-1254-4	(4)	7.26	7.25	7.15	7.35	-0.01
Aroclor-1254-5	(5)	7.68	7.68	7.58	7.78	0.00
Tetrachloro-m-xylene		4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl		10.09	10.09	9.99	10.19	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

Continuing Calib Date: 11/19/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 20:53 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.52	5.52	5.42	5.62	0.00
Aroclor-1254-2 (2)	5.67	5.67	5.57	5.77	0.00
Aroclor-1254-3 (3)	6.07	6.07	5.97	6.17	0.00
Aroclor-1254-4 (4)	6.30	6.30	6.20	6.40	0.00
Aroclor-1254-5 (5)	6.72	6.72	6.62	6.82	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.63	8.63	8.53	8.73	0.00