

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL01 Date Analyzed: 11/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO108090.D Time Analyzed: 09:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.527	5.425	5.625	517.960	500.000	3.6
Aroclor-1016-2	5.549	5.447	5.647	518.210	500.000	3.6
Aroclor-1016-3	5.612	5.510	5.710	508.200	500.000	1.6
Aroclor-1016-4	5.709	5.607	5.807	522.050	500.000	4.4
Aroclor-1016-5	6.005	5.904	6.104	523.510	500.000	4.7
Aroclor-1260-1	7.139	7.037	7.237	519.840	500.000	4.0
Aroclor-1260-2	7.396	7.295	7.495	548.610	500.000	9.7
Aroclor-1260-3	7.760	7.658	7.858	530.810	500.000	6.2
Aroclor-1260-4	7.985	7.884	8.084	543.060	500.000	8.6
Aroclor-1260-5	8.300	8.198	8.398	545.640	500.000	9.1
Decachlorobiphenyl	10.095	9.991	10.191	53.260	50.000	6.5
Tetrachloro-m-xylene	4.372	4.270	4.470	53.000	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL01 Date Analyzed: 11/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO108090.D Time Analyzed: 09:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.620	4.820	514.990	500.000	3.0
Aroclor-1016-2	4.739	4.639	4.839	523.120	500.000	4.6
Aroclor-1016-3	4.915	4.815	5.015	524.490	500.000	4.9
Aroclor-1016-4	4.957	4.856	5.056	517.640	500.000	3.5
Aroclor-1016-5	5.170	5.070	5.270	525.250	500.000	5.1
Aroclor-1260-1	6.201	6.100	6.300	517.640	500.000	3.5
Aroclor-1260-2	6.389	6.288	6.488	519.540	500.000	3.9
Aroclor-1260-3	6.542	6.441	6.641	516.100	500.000	3.2
Aroclor-1260-4	7.013	6.912	7.112	518.930	500.000	3.8
Aroclor-1260-5	7.254	7.154	7.354	520.740	500.000	4.1
Decachlorobiphenyl	8.634	8.535	8.735	50.730	50.000	1.5
Tetrachloro-m-xylene	3.639	3.539	3.739	52.510	50.000	5.0