

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL02 Date Analyzed: 11/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO108103.D Time Analyzed: 16:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.524	5.425	5.625	515.170	500.000	3.0
Aroclor-1016-2	5.547	5.447	5.647	520.530	500.000	4.1
Aroclor-1016-3	5.610	5.510	5.710	508.600	500.000	1.7
Aroclor-1016-4	5.707	5.607	5.807	522.040	500.000	4.4
Aroclor-1016-5	6.003	5.904	6.104	530.110	500.000	6.0
Aroclor-1260-1	7.136	7.037	7.237	525.910	500.000	5.2
Aroclor-1260-2	7.394	7.295	7.495	559.580	500.000	11.9
Aroclor-1260-3	7.756	7.658	7.858	539.160	500.000	7.8
Aroclor-1260-4	7.982	7.884	8.084	548.890	500.000	9.8
Aroclor-1260-5	8.298	8.198	8.398	552.520	500.000	10.5
Decachlorobiphenyl	10.092	9.991	10.191	53.750	50.000	7.5
Tetrachloro-m-xylene	4.370	4.270	4.470	52.920	50.000	5.8

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL02 Date Analyzed: 11/19/2024

Lab Sample No.: AR1660CCC500 Data File : PO108103.D Time Analyzed: 16:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.620	4.820	512.440	500.000	2.5
Aroclor-1016-2	4.737	4.639	4.839	521.260	500.000	4.3
Aroclor-1016-3	4.914	4.815	5.015	520.400	500.000	4.1
Aroclor-1016-4	4.955	4.856	5.056	515.180	500.000	3.0
Aroclor-1016-5	5.168	5.070	5.270	519.470	500.000	3.9
Aroclor-1260-1	6.199	6.100	6.300	517.700	500.000	3.5
Aroclor-1260-2	6.387	6.288	6.488	515.130	500.000	3.0
Aroclor-1260-3	6.541	6.441	6.641	512.730	500.000	2.5
Aroclor-1260-4	7.012	6.912	7.112	513.710	500.000	2.7
Aroclor-1260-5	7.253	7.154	7.354	513.270	500.000	2.7
Decachlorobiphenyl	8.633	8.535	8.735	49.900	50.000	-0.2
Tetrachloro-m-xylene	3.637	3.539	3.739	52.380	50.000	4.8