

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL03 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108119.D Time Analyzed: 09:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.528	5.425	5.625	491.220	500.000	-1.8
Aroclor-1016-2	5.550	5.447	5.647	494.320	500.000	-1.1
Aroclor-1016-3	5.612	5.510	5.710	476.060	500.000	-4.8
Aroclor-1016-4	5.710	5.607	5.807	483.320	500.000	-3.3
Aroclor-1016-5	6.006	5.904	6.104	495.840	500.000	-0.8
Aroclor-1260-1	7.141	7.037	7.237	511.610	500.000	2.3
Aroclor-1260-2	7.399	7.295	7.495	544.230	500.000	8.8
Aroclor-1260-3	7.762	7.658	7.858	518.550	500.000	3.7
Aroclor-1260-4	7.987	7.884	8.084	530.680	500.000	6.1
Aroclor-1260-5	8.303	8.198	8.398	539.820	500.000	8.0
Decachlorobiphenyl	10.098	9.991	10.191	53.340	50.000	6.7
Tetrachloro-m-xylene	4.372	4.270	4.470	51.090	50.000	2.2

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL03 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108119.D Time Analyzed: 09:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.620	4.820	523.940	500.000	4.8
Aroclor-1016-2	4.738	4.639	4.839	527.440	500.000	5.5
Aroclor-1016-3	4.914	4.815	5.015	528.130	500.000	5.6
Aroclor-1016-4	4.956	4.856	5.056	517.930	500.000	3.6
Aroclor-1016-5	5.169	5.070	5.270	531.540	500.000	6.3
Aroclor-1260-1	6.201	6.100	6.300	525.860	500.000	5.2
Aroclor-1260-2	6.389	6.288	6.488	519.470	500.000	3.9
Aroclor-1260-3	6.542	6.441	6.641	518.410	500.000	3.7
Aroclor-1260-4	7.013	6.912	7.112	518.810	500.000	3.8
Aroclor-1260-5	7.254	7.154	7.354	519.750	500.000	4.0
Decachlorobiphenyl	8.635	8.535	8.735	50.790	50.000	1.6
Tetrachloro-m-xylene	3.638	3.539	3.739	53.620	50.000	7.2