

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL04 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108134.D Time Analyzed: 14:51

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|----------------------|--------|-----------|--------|--------------------|-------------------|-----|
|                      |        | FROM      | TO     |                    |                   |     |
| Aroclor-1016-1       | 5.525  | 5.425     | 5.625  | 521.430            | 500.000           | 4.3 |
| Aroclor-1016-2       | 5.548  | 5.447     | 5.647  | 525.660            | 500.000           | 5.1 |
| Aroclor-1016-3       | 5.610  | 5.510     | 5.710  | 517.460            | 500.000           | 3.5 |
| Aroclor-1016-4       | 5.708  | 5.607     | 5.807  | 532.790            | 500.000           | 6.6 |
| Aroclor-1016-5       | 6.004  | 5.904     | 6.104  | 530.560            | 500.000           | 6.1 |
| Aroclor-1260-1       | 7.138  | 7.037     | 7.237  | 511.560            | 500.000           | 2.3 |
| Aroclor-1260-2       | 7.395  | 7.295     | 7.495  | 542.950            | 500.000           | 8.6 |
| Aroclor-1260-3       | 7.759  | 7.658     | 7.858  | 527.760            | 500.000           | 5.6 |
| Aroclor-1260-4       | 7.984  | 7.884     | 8.084  | 540.130            | 500.000           | 8.0 |
| Aroclor-1260-5       | 8.300  | 8.198     | 8.398  | 540.760            | 500.000           | 8.2 |
| Decachlorobiphenyl   | 10.093 | 9.991     | 10.191 | 52.860             | 50.000            | 5.7 |
| Tetrachloro-m-xylene | 4.371  | 4.270     | 4.470  | 53.890             | 50.000            | 7.8 |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL04 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108134.D Time Analyzed: 14:51

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.719 | 4.620                | 4.820 | 531.180            | 500.000           | 6.2  |
| Aroclor-1016-2       | 4.738 | 4.639                | 4.839 | 535.030            | 500.000           | 7.0  |
| Aroclor-1016-3       | 4.914 | 4.815                | 5.015 | 538.150            | 500.000           | 7.6  |
| Aroclor-1016-4       | 4.956 | 4.856                | 5.056 | 526.910            | 500.000           | 5.4  |
| Aroclor-1016-5       | 5.169 | 5.070                | 5.270 | 537.190            | 500.000           | 7.4  |
| Aroclor-1260-1       | 6.200 | 6.100                | 6.300 | 526.660            | 500.000           | 5.3  |
| Aroclor-1260-2       | 6.388 | 6.288                | 6.488 | 519.760            | 500.000           | 4.0  |
| Aroclor-1260-3       | 6.540 | 6.441                | 6.641 | 510.520            | 500.000           | 2.1  |
| Aroclor-1260-4       | 7.012 | 6.912                | 7.112 | 510.320            | 500.000           | 2.1  |
| Aroclor-1260-5       | 7.253 | 7.154                | 7.354 | 508.310            | 500.000           | 1.7  |
| Decachlorobiphenyl   | 8.634 | 8.535                | 8.735 | 48.500             | 50.000            | -3.0 |
| Tetrachloro-m-xylene | 3.638 | 3.539                | 3.739 | 54.200             | 50.000            | 8.4  |