

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL06 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108159.D Time Analyzed: 23:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.526	5.425	5.625	545.410	500.000	9.1
Aroclor-1016-2	5.548	5.447	5.647	539.240	500.000	7.8
Aroclor-1016-3	5.611	5.510	5.710	531.900	500.000	6.4
Aroclor-1016-4	5.708	5.607	5.807	546.420	500.000	9.3
Aroclor-1016-5	6.005	5.904	6.104	552.530	500.000	10.5
Aroclor-1260-1	7.138	7.037	7.237	546.400	500.000	9.3
Aroclor-1260-2	7.397	7.295	7.495	577.560	500.000	15.5
Aroclor-1260-3	7.760	7.658	7.858	556.610	500.000	11.3
Aroclor-1260-4	7.985	7.884	8.084	559.850	500.000	12.0
Aroclor-1260-5	8.301	8.198	8.398	566.220	500.000	13.2
Decachlorobiphenyl	10.097	9.991	10.191	53.910	50.000	7.8
Tetrachloro-m-xylene	4.371	4.270	4.470	56.130	50.000	12.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL06 Date Analyzed: 11/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108159.D Time Analyzed: 23:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.620	4.820	547.280	500.000	9.5
Aroclor-1016-2	4.738	4.639	4.839	549.370	500.000	9.9
Aroclor-1016-3	4.914	4.815	5.015	546.590	500.000	9.3
Aroclor-1016-4	4.956	4.856	5.056	532.480	500.000	6.5
Aroclor-1016-5	5.169	5.070	5.270	557.300	500.000	11.5
Aroclor-1260-1	6.200	6.100	6.300	549.930	500.000	10.0
Aroclor-1260-2	6.388	6.288	6.488	546.610	500.000	9.3
Aroclor-1260-3	6.541	6.441	6.641	538.450	500.000	7.7
Aroclor-1260-4	7.013	6.912	7.112	537.050	500.000	7.4
Aroclor-1260-5	7.254	7.154	7.354	536.090	500.000	7.2
Decachlorobiphenyl	8.636	8.535	8.735	50.810	50.000	1.6
Tetrachloro-m-xylene	3.638	3.539	3.739	56.020	50.000	12.0