

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL01 Date Analyzed: 11/22/2024

Lab Sample No.: HSTDCCC750 Data File : PS028596.D Time Analyzed: 10:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.236	9.150	9.350	725.540	712.500	1.8
2,4-D	8.358	8.270	8.470	706.690	705.000	0.2
2,4-DCAA	7.232	7.141	7.341	743.190	750.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL01 Date Analyzed: 11/22/2024

Lab Sample No.: HSTDCCC750 Data File : PS028596.D Time Analyzed: 10:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.883	9.796	9.996	676.920	712.500	-5.0
2,4-D	8.978	8.890	9.090	645.580	705.000	-8.4
2,4-DCAA	7.732	7.642	7.842	661.550	750.000	-11.8