

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL02 Date Analyzed: 11/22/2024

Lab Sample No.: HSTDCCC750 Data File : PS028604.D Time Analyzed: 14:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.234	9.150	9.350	747.520	712.500	4.9
2,4-D	8.357	8.270	8.470	727.490	705.000	3.2
2,4-DCAA	7.231	7.141	7.341	766.810	750.000	2.2

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL02 Date Analyzed: 11/22/2024

Lab Sample No.: HSTDCCC750 Data File : PS028604.D Time Analyzed: 14:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.881	9.796	9.996	710.800	712.500	-0.2
2,4-D	8.976	8.890	9.090	679.390	705.000	-3.6
2,4-DCAA	7.731	7.642	7.842	688.910	750.000	-8.1