

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL03 Date Analyzed: 11/22/2024

Lab Sample No.: PSTDCCC050 Data File : PL093215.D Time Analyzed: 10:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.062	8.954	9.154	44.180	50.000	-11.6
Endrin	6.575	6.474	6.674	42.370	50.000	-15.3
gamma-BHC (Lindane)	4.327	4.228	4.428	48.930	50.000	-2.1
Heptachlor	4.916	4.817	5.017	46.560	50.000	-6.9
Heptachlor epoxide	5.685	5.584	5.784	46.450	50.000	-7.1
Methoxychlor	7.503	7.399	7.599	41.130	50.000	-17.7
Tetrachloro-m-xylene	3.538	3.440	3.640	49.870	50.000	-0.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL03 Date Analyzed: 11/22/2024

Lab Sample No.: PSTDCCC050 Data File : PL093215.D Time Analyzed: 10:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.917	7.816	8.016	49.590	50.000	-0.8
Endrin	5.640	5.541	5.741	49.080	50.000	-1.8
gamma-BHC (Lindane)	3.608	3.511	3.711	51.140	50.000	2.3
Heptachlor	3.947	3.849	4.049	49.890	50.000	-0.2
Heptachlor epoxide	4.730	4.632	4.832	51.160	50.000	2.3
Methoxychlor	6.615	6.515	6.715	47.330	50.000	-5.3
Tetrachloro-m-xylene	2.775	2.678	2.878	49.800	50.000	-0.4