

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL04 Date Analyzed: 11/22/2024

Lab Sample No.: PSTDCCC050 Data File : PL093228.D Time Analyzed: 16:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.063	8.954	9.154	43.200	50.000	-13.6
Endrin	6.575	6.474	6.674	47.400	50.000	-5.2
gamma-BHC (Lindane)	4.328	4.228	4.428	50.720	50.000	1.4
Heptachlor	4.917	4.817	5.017	49.570	50.000	-0.9
Heptachlor epoxide	5.684	5.584	5.784	47.450	50.000	-5.1
Methoxychlor	7.504	7.399	7.599	45.380	50.000	-9.2
Tetrachloro-m-xylene	3.539	3.440	3.640	50.980	50.000	2.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL04 Date Analyzed: 11/22/2024

Lab Sample No.: PSTDCCC050 Data File : PL093228.D Time Analyzed: 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.917	7.816	8.016	49.440	50.000	-1.1
Endrin	5.639	5.541	5.741	54.970	50.000	9.9
gamma-BHC (Lindane)	3.607	3.511	3.711	52.030	50.000	4.1
Heptachlor	3.946	3.849	4.049	52.950	50.000	5.9
Heptachlor epoxide	4.729	4.632	4.832	53.710	50.000	7.4
Methoxychlor	6.614	6.515	6.715	53.380	50.000	6.8
Tetrachloro-m-xylene	2.774	2.678	2.878	50.510	50.000	1.0