

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: P4892 SDG NO.: P4892

Lab File ID: BF140331.D

DFTPP Injection Date: 11/13/2024

Instrument ID: BNA_F

DFTPP Injection Time: 08:35

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	36.9
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	38
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	48.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
275	10.0 - 60.0% of mass 198	28.3
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	14
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.7 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF140332.D	11/13/2024	09:01
SSTDICC005	SSTDICC005	BF140333.D	11/13/2024	09:27
SSTDICC010	SSTDICC010	BF140334.D	11/13/2024	09:53
SSTDICC020	SSTDICC020	BF140335.D	11/13/2024	10:29
SSTDICC050	SSTDICC050	BF140337.D	11/13/2024	11:21
SSTDICC060	SSTDICC060	BF140338.D	11/13/2024	11:47
SSTDICC080	SSTDICC080	BF140339.D	11/13/2024	12:13
SSTDICCC040	SSTDICCC040	BF140340.D	11/13/2024	12:48