

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112724\
 Data File : BF140655.D
 Acq On : 27 Nov 2024 08:47
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 27 10:18:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	40.309	-0.8	93	-0.02
3	Pyridine	40.000	42.993	-7.5	89	-0.02
4	n-Nitrosodimethylamine	40.000	37.353	6.6	84	-0.02
5 S	2-Fluorophenol	80.000	77.683	2.9	89	0.00
6	Aniline	40.000	47.062	-17.7	94	0.00
7 S	Phenol-d6	80.000	77.515	3.1	86	0.00
8	2-Chlorophenol	40.000	39.359	1.6	88	0.00
9	Benzaldehyde	40.000	36.487	8.8	93	0.00
10 C	Phenol	40.000	40.043	-0.1	87	0.00
11	bis(2-Chloroethyl)ether	40.000	39.060	2.3	87	0.00
12	1,3-Dichlorobenzene	40.000	39.038	2.4	90	0.00
13 C	1,4-Dichlorobenzene	40.000	39.341	1.6	90	0.00
14	1,2-Dichlorobenzene	40.000	39.113	2.2	89	0.00
15	Benzyl Alcohol	40.000	39.178	2.1	84	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.299	1.8	88	0.00
17	2-Methylphenol	40.000	39.213	2.0	87	0.00
18	Hexachloroethane	40.000	39.205	2.0	89	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.390	6.5	83	0.00
20	3+4-Methylphenols	40.000	38.790	3.0	85	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	88	0.00
22	Acetophenone	40.000	38.154	4.6	85	0.00
23 S	Nitrobenzene-d5	80.000	77.548	3.1	85	0.00
24	Nitrobenzene	40.000	38.398	4.0	85	0.00
25	Isophorone	40.000	38.477	3.8	83	0.00
26 C	2-Nitrophenol	40.000	39.751	0.6	87	0.00
27	2,4-Dimethylphenol	40.000	41.498	-3.7	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.189	2.0	86	0.00
29 C	2,4-Dichlorophenol	40.000	39.535	1.2	87	0.00
30	1,2,4-Trichlorobenzene	40.000	38.749	3.1	87	0.00
31	Naphthalene	40.000	38.933	2.7	87	0.00
32	Benzoic acid	40.000	37.340	6.6	82	0.00
33	4-Chloroaniline	40.000	43.304	-8.3	90	0.00
34 C	Hexachlorobutadiene	40.000	39.137	2.2	87	0.00
35	Caprolactam	40.000	38.477	3.8	82	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.156	2.1	83	0.00
37	2-Methylnaphthalene	40.000	38.836	2.9	86	0.00
38	1-Methylnaphthalene	40.000	39.041	2.4	86	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.146	2.1	86	0.00
41 P	Hexachlorocyclopentadiene	40.000	31.653	20.9	65	0.00
42 S	2,4,6-Tribromophenol	80.000	77.572	3.0	81	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.498	3.8	82	0.00
44	2,4,5-Trichlorophenol	40.000	39.720	0.7	83	0.00
45 S	2-Fluorobiphenyl	80.000	77.473	3.2	85	0.00
46	1,1'-Biphenyl	40.000	39.347	1.6	86	0.00
47	2-Chloronaphthalene	40.000	39.742	0.6	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112724\
 Data File : BF140655.D
 Acq On : 27 Nov 2024 08:47
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 27 10:18:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.925	2.7	81	0.00
49	Acenaphthylene	40.000	39.584	1.0	86	0.00
50	Dimethylphthalate	40.000	39.265	1.8	84	0.00
51	2,6-Dinitrotoluene	40.000	39.422	1.4	83	0.00
52 C	Acenaphthene	40.000	39.008	2.5	84	0.00
53	3-Nitroaniline	40.000	41.743	-4.4	86	0.00
54 P	2,4-Dinitrophenol	40.000	38.107	4.7	79	0.00
55	Dibenzofuran	40.000	38.450	3.9	83	0.00
56 P	4-Nitrophenol	40.000	37.175	7.1	76	0.00
57	2,4-Dinitrotoluene	40.000	40.059	-0.1	83	0.00
58	Fluorene	40.000	38.840	2.9	84	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.168	-0.4	83	0.00
60	Diethylphthalate	40.000	37.864	5.3	82	0.00
61	4-Chlorophenyl-phenylether	40.000	38.606	3.5	83	0.00
62	4-Nitroaniline	40.000	40.295	-0.7	84	0.00
63	Azobenzene	40.000	38.236	4.4	83	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	85	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.583	-9.0	86	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.535	3.7	84	0.00
67	4-Bromophenyl-phenylether	40.000	38.901	2.7	85	0.00
68	Hexachlorobenzene	40.000	39.247	1.9	85	0.00
69	Atrazine	40.000	37.801	5.5	96	0.00
70 C	Pentachlorophenol	40.000	38.787	3.0	75	0.00
71	Phenanthrene	40.000	38.970	2.6	84	0.00
72	Anthracene	40.000	39.499	1.3	85	0.00
73	Carbazole	40.000	40.646	-1.6	88	0.00
74	Di-n-butylphthalate	40.000	38.654	3.4	84	0.00
75 C	Fluoranthene	40.000	41.207	-3.0	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	41.752	-4.4	103	0.00
78	Pyrene	40.000	37.535	6.2	91	0.00
79 S	Terphenyl-d14	80.000	72.402	9.5	89	0.00
80	Butylbenzylphthalate	40.000	38.481	3.8	92	0.00
81	Benzo(a)anthracene	40.000	39.247	1.9	99	0.00
82	3,3'-Dichlorobenzidine	40.000	38.833	2.9	91	0.00
83	Chrysene	40.000	39.911	0.2	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.045	-0.1	99	0.00
85 c	Di-n-octyl phthalate	40.000	41.221	-3.1	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.145	4.6	84	0.01
88	Benzo(b)fluoranthene	40.000	43.241	-8.1	100	0.00
89	Benzo(k)fluoranthene	40.000	38.282	4.3	84	0.00
90 C	Benzo(a)pyrene	40.000	39.968	0.1	89	0.01
91	Dibenzo(a,h)anthracene	40.000	37.832	5.4	83	0.01
92	Benzo(g,h,i)perylene	40.000	38.588	3.5	85	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112724\
Data File : BF140655.D
Acq On : 27 Nov 2024 08:47
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 27 10:18:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0