

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

EPA Sample No.: SSTDICCC040 Date Analyzed: 11/21/2024

Lab File ID: BF140532.D Time Analyzed: 12:58

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	107516	6.875	413408	8.16	234407	9.92
UPPER LIMIT	215032	7.375	826816	8.657	468814	10.416
LOWER LIMIT	53758	6.375	206704	7.657	117204	9.416
EPA SAMPLE NO.						
01 PB165144BL	89214	6.87	332852	8.15	187896	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.