

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/21/2024

Lab File ID: BF140539.D Time Analyzed: 16:54

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	106707	6.875	411509	8.16	234805	9.91
UPPER LIMIT	213414	7.375	823018	8.657	469610	10.41
LOWER LIMIT	53353.5	6.375	205755	7.657	117403	9.41
EPA SAMPLE NO.						
01 PB165060TB	89751	6.87	336762	8.15	191525	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.