

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892

EPA Sample No.: SSTDICCC040 Date Analyzed: 11/21/2024

Lab File ID: BF140532.D Time Analyzed: 12:58

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	437466	11.404	239343	14.051	211422	15.539
UPPER LIMIT	874932	11.904	478686	14.551	422844	16.039
LOWER LIMIT	218733	10.904	119672	13.551	105711	15.039
EPA SAMPLE NO.						
01 PB165144BL	359986	11.40	208110	14.05	161923	15.54

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.