

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: P4892 SDG NO.: P4892

Lab File ID: BF140526.D

DFTPP Injection Date: 11/21/2024

Instrument ID: BNA_F

DFTPP Injection Time: 10:17

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.3
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	35.5
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	48
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	29.2
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	15.7
442	Greater than 50% of mass 198	99.8
443	15.0 - 24.0% of mass 442	18.1 (18.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF140528.D	11/21/2024	11:13
SSTDICC005	SSTDICC005	BF140529.D	11/21/2024	11:39
SSTDICC010	SSTDICC010	BF140530.D	11/21/2024	12:05
SSTDICC020	SSTDICC020	BF140531.D	11/21/2024	12:32
SSTDICCC040	SSTDICCC040	BF140532.D	11/21/2024	12:58
SSTDICC050	SSTDICC050	BF140533.D	11/21/2024	13:25
SSTDICC060	SSTDICC060	BF140534.D	11/21/2024	13:51
SSTDICC080	SSTDICC080	BF140535.D	11/21/2024	14:18
PB165144BL	PB165144BL	BF140537.D	11/21/2024	15:34