

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG No.: P4892

Instrument ID: MSVOA_N Calibration Date/Time: 11/21/2024 15:22

Lab File ID: VN084995.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.618	0.584		-5.5	20
1,1-Dichloroethene	0.569	0.574		0.88	20
2-Butanone	0.337	0.448		32.94	20
Carbon Tetrachloride	0.525	0.606		15.43	20
Chloroform	1.121	1.295		15.52	20
Benzene	1.507	1.618		7.37	20
1,2-Dichloroethane	0.488	0.558		14.34	20
Trichloroethene	0.348	0.375		7.76	20
Tetrachloroethene	0.333	0.360		8.11	20
Chlorobenzene	1.119	1.238	0.3	10.63	20
1,2-Dichloroethane-d4	0.722	0.673		-6.79	20
Dibromofluoromethane	0.338	0.331		-2.07	20
Toluene-d8	1.247	1.154		-7.46	20
4-Bromofluorobenzene	0.466	0.442		-5.15	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.