

Surrogate Summary

SDG No.: **P4892**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4892-03	WB-310-BOT	1,2-Dichloroethane-d4	50	51.3	102	70 (74)	130 (125)
		Dibromofluoromethane	50	47.6	95	70 (75)	130 (124)
		Toluene-d8	50	46.7	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.2	92	70 (77)	130 (121)
VN1121WBL02	VN1121WBL02	1,2-Dichloroethane-d4	50	50.5	101	70 (74)	130 (125)
		Dibromofluoromethane	50	48.6	97	70 (75)	130 (124)
		Toluene-d8	50	46.1	92	70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.9	88	70 (77)	130 (121)
VN1121WBS02	VN1121WBS02	1,2-Dichloroethane-d4	50	49.8	100	70 (74)	130 (125)
		Dibromofluoromethane	50	51.1	102	70 (75)	130 (124)
		Toluene-d8	50	47.6	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.9	98	70 (77)	130 (121)
VN1121WBSD0	VN1121WBSD02	1,2-Dichloroethane-d4	50	52.5	105	70 (74)	130 (125)
		Dibromofluoromethane	50	51.4	103	70 (75)	130 (124)
		Toluene-d8	50	47.5	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.3	103	70 (77)	130 (121)

