

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4892

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VN084999.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1121WBSD02	Vinyl chloride	20	17.1	ug/L	86	10		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	17.6	ug/L	88	12		70 (74)	130 (110)	20 (20)
	2-Butanone	100	130	ug/L	130	17		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	19.9	ug/L	100	5		70 (77)	130 (113)	20 (20)
	Chloroform	20	21.4	ug/L	107	13		70 (79)	130 (113)	20 (20)
	Benzene	20	19.4	ug/L	97	9		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	21.6	ug/L	108	10		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.1	ug/L	96	8		70 (77)	130 (113)	20 (20)
	Tetrachloroethene	20	18.8	ug/L	94	7		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	19.5	ug/L	98	9		70 (82)	130 (109)	20 (20)