

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892
Lab File ID: VN084995.D Date Analyzed: 11/21/2024
Instrument ID: MSVOA_N Time Analyzed: 15:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	156906	8.22	261929	9.10	230917	11.87
UPPER LIMIT	313812	8.718	523858	9.6	461834	12.365
LOWER LIMIT	78453	7.718	130965	8.6	115459	11.365
EPA SAMPLE NO.						
WB-310-BOT	163820	8.22	297677	9.10	265873	11.87
VN1121WBL02	178139	8.22	314864	9.10	264684	11.87
VN1121WBS02	154131	8.22	258963	9.09	225356	11.87
VN1121WBSD02	138009	8.22	233557	9.10	207781	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.