

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892  
 Lab File ID: VN084995.D Date Analyzed: 11/21/2024  
 Instrument ID: MSVOA\_N Time Analyzed: 15:22  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 117373        | 13.788 |  |  |  |  |
| UPPER LIMIT    | 234746        | 14.288 |  |  |  |  |
| LOWER LIMIT    | 58686.5       | 13.288 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| WB-310-BOT     | 111856        | 13.79  |  |  |  |  |
| VN1121WBL02    | 112398        | 13.79  |  |  |  |  |
| VN1121WBS02    | 114991        | 13.79  |  |  |  |  |
| VN1121WBSD02   | 103882        | 13.79  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.