

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4892 SAS No.: P4892 SDG NO.: P4892
 Lab File ID: VN084960.D Date Analyzed: 11/20/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:18
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	164992	8.22	271316	9.10	240365	11.87
UPPER LIMIT	329984	8.724	542632	9.6	480730	12.365
LOWER LIMIT	82496	7.724	135658	8.6	120183	11.365
EPA SAMPLE NO.						
WB-310-SW	166877	8.22	307613	9.10	262227	11.87
VN1120WBL01	178908	8.22	319489	9.10	277659	11.87
VN1120WBS02	150630	8.22	253557	9.10	229692	11.87
VN1120WBSD02	152907	8.22	266547	9.10	231842	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.